

AMINES

Column: **CGB-624**, P/N CG-603065

Size: 60m x 0.53mm x 3.0 μ m

Injection: 1 μ l amines standard, split 1:5, 260 $^{\circ}$ C

Carrier Gas: He, 8 mL/min

Program temperature: 40 $^{\circ}$ C (1min) @ 5 $^{\circ}$ C/min to 260 $^{\circ}$ C (10min)

Detector: FID KONIK-TECH, 270 $^{\circ}$ C

Sample

PM (1-methoxy-2-propanol)

DMEA (N,N-dimethylethanolamine)

BDMA (N,N-dimethylbenzylamine)

PMDTA (pentamethyldiethylentriamine)

TDMAPA (N,N,N-tris(3-dimethylaminopropyl)amine)

Chromatogram provided by Ariadna Galve from KONIK-TECH, S.A

