

ANALYSIS OF POLYOLS IN WINE

Column: **CGB-FFAP**, P/N CG-150262

Dimensions: 60m x 0.25mm x 0.25 μ m

Injection: 1 μ L, split (100:1), glycols standard, 205°C

Carrier gas: H₂, 1 mL/min (80°C)

Oven temperature: 100°C @ 5°C/min to 200°C(10 min)

Detector: FID, 275°C

Chromatogram provided by R. Franquet and J. García from INCAVI, Vilafranca del Penedès (Barcelona)

